

## FEATURES

- Fingertip Textured
- Powder Free
- Not made with natural rubber latex
- Chemo drugs tested
- Lab chemical tested
- Single use only
- Ambidextrous
- Patient examination glove

## SPECIFICATIONS

Type	Powder free and non sterile
Material	100% nitrile (Acrylonitrile-butadiene). No natural rubber latex
Design	Ambidextrous, beaded cuff and smooth micro textured finish for easier grip
Colour	Blue
Size	X Small, Small, Medium, Large, X Large
Donning Agent	Powder free glove
Storage & Shelf Life	The gloves shall maintain their properties when stored in a dry condition at temperature between 5 to 40°C. Protect gloves against ultra violet light sources, such as sunlight and oxidizing agents. Copper ions discolour the glove. 5 Years from the date of manufacturing.

### Chemotherapy Drugs and Concentration

(Tested for resistance to Permeation by Chemotherapy Drugs according to ASTM D6978 test methods and requirements)

### Minimum Breakthrough Detection Time

Cisplatin, 1.0mg/ml	>240 Minutes
Cyclophosphamide (Cytoxan), 20.0mg/ml	>240 Minutes
Dacarbazine (DTIC), 10.0mg/ml	>240 Minutes
Doxorubicin Hydrochloride, 2.0mg/ml	>240 Minutes
Etoposide (Toposar), 20.00mg/ml	>240 Minutes
Fluorouracil, 50.0mg/ml	>240 Minutes
Mitomycin C, 0.5 mg/ml	>240 Minutes
Paclitaxel (Taxol), 6.0mg/ml	>240 Minutes
Vincristine Sulfate, 1 mg/ml	>240 Minutes

### Chemical

### EN 16523-1 Permeation Level

### EN 374-4 Mean Degradation (%)

37% Formaldehyde (CAS: 50-00-0)	Level 3	0.5
40% Sodium Hydroxide (CAS: 1310-73-2)	Level 6	-39.2
30% Hydrogen Peroxide (CAS: 7722-84-1)	Level 2	30.8

### Measured Breakthrough Time (Minutes)

>10   >30   >60   >120   >240   >480

### Permeation Performance Level

1   2   3   4   5   6



Part No: **GTD10**

## SIZING

X Small	GTD100
Small	GTD101
Medium	GTD102
Large	GTD103
X Large	GTD104

## PACKAGING

200 gloves per box for XS to L  
180 gloves per box for XL  
10 boxes per carton

## REGULATORY COMPLIANCE

ARTG 438772  
US FDA 21-CFR 177  
REACH Certified  
HACCP Certified  
EU CE Regulation 2016/425

## STANDARDS

ASTM D6978  
EN 16523-1  
EN 420  
EN 374 (1;4;5)  
EN 455 (1;2;3;4)  
EN 1186  
ISO 10993.10:2016

## MANUFACTURING CERTIFICATIONS

ISO 9001:2015	International Quality Management System
ISO 13485	Medical Devices Quality Management System
ISO 14001	Production & Distribution of Examination Gloves

### STANDARDS & CERTIFICATIONS



Resistance to Bacteria, Fungi and Virus.



COMPLIANCE: AS/NZS 4011.1:2014

